



406 Dividend Drive
 Peachtree City, Georgia 30269
 Phone (770) 632-1851
 Fax: (770) 632-1852

Repair Estimate

	YES	NO
LOG CARD:	<input type="checkbox"/>	<input type="checkbox"/>
SPECIAL INSTRUCTION:	<input type="checkbox"/>	<input type="checkbox"/>
UNIT MISUSED:	<input type="checkbox"/>	<input checked="" type="checkbox"/>
PICTURES:	<input type="checkbox"/>	<input checked="" type="checkbox"/>

DATE:	4-May-2022
PRIME REFERENCE:	051184
CUSTOMER REFERENCE:	RMN-22-320
PART NUMBER:	5677-771-00-10
SERIAL NUMBER:	458
DESCRIPTION:	TORQUEMETER INDICATOR
WORK REQUESTED:	REPAIRED

EVALUATIONS / FINDINGS:

TESTED UNIT. COULD NOT DUPLICATE ERRATIC READINGS. FOUND #2 CALIBRATION POTS GROMMETS HAVE BEEN PUSH INTO CASE AND ZERO POT IS ADJUSTED TO THE EXTREME. #2 POINTER IS WAY OUT OF CALIBRATION
 INSPECTED ALL CIRCUIT BOARDS NO ISSUE FOUND.

CORRECTIVE ACTIONS TO BE CARRIED OUT:

REPAIRED REQUIRED

PARTS REPLACED OR REPAIRED

EXCLUSIONS:	NO	CONDITION AFTER COMPLETION:	REPAIRED
-------------	----	-----------------------------	----------

PART NUMBER	DESCRIPTION	STATUS	QUANTITY	PRICE

FLAT RATE:	\$795.00	COMMENTS: WARRANTY DENIED
TOTAL QUOTE		
TOTAL QUOTE OPTION #1		
TURN AROUND TIME AFTER APPROVAL: (IN BUSINESS DAYS)	30	
EVALUATION FEE IF QUOTE DENIED	\$500.00	

Approval Signature: _____
 Name: _____ Date: _____