



SMSB

SENDI MAHIR SDN. BHD. 199501003943 (333138-T)

NO. 6, 8, 10 & 12, JALAN KAPAR 27/89, MEGAH INDUSTRIAL PARK,
SEKSYEN 27, 40400 SHAH ALAM, SELANGOR DARUL EHSAN, MALAYSIA.
TEL: 03-5191 7388 FAX: 03-5191 0675

EMAIL: enquiry@sendimahir.com ; marketing@sendimahir.com Website: www.sendimahir.com



SAMM 082

CERTIFICATE OF CALIBRATION

Certificate No. : SM24111487
Issued By : Sendi Mahir Sdn Bhd

Date of Issue : 14 Feb 2024

Page 1 of 3 Pages

SM24111487-624

Customer : GALAXY AEROSPACE (M) SDN BHD
SUITE 11-14, HELICOPTER CENTRE
MALAYSIA INTERNATIONAL AEROSPACE CENTRE SULTAN ABDUL AZIZ SHAH AIRPORT
47200 SUBANG SELANGOR

Instrument : Automatic Reflex Battery Charger/Analyzer
Manufacturer : Christie
Model/Type : RF80-M
Serial No : 7240 (CTE/250)
Capacity : Refer to Following Pages
Resolution : -

Calibration Date : 13 Feb 2024
Recalibration Date Specified By Customer : 13 Feb 2025

Condition Upon Receiving : Good Physical Condition
Condition Upon Returning : The instrument has been calibrated. The results are as follows. No adjustment done.

Calibration Method : Internal Calibration Procedure(s) ICPE 1/18
Calibration Venue : This Instrument has been calibrated at Company as above

Measurement Uncertainty : The reported expanded measurement uncertainty is stated as the standard measurement uncertainty multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95% and have a coverage factor of k=2 unless stated otherwise.

Remark : The user should be aware that any numbers of factors may cause this instrument to drift out of calibration before the specified calibration interval has expired.

Calibration Environment Condition:
Temperature : 21 to 22 °C
Relative Humidity : 55 to 60 %rh

Reference Standard(s) Used :

Reference Standard Name	Serial No	Certificate No	Due Date	Accreditation No	Traceability
PRECISION MULTIMETER	E003	SM24102804	23 Jan 2025	SAMM 082	NMIM(MY)
AC/DC ACTIVE CURRENT SHUNT	E059	NMIM-5191-E-23	09 Nov 2025	SAMM 261	NMIM(MY)
PROCESS CALIBRATOR	TS079	SM24107042	31 Jan 2025	SAMM 082	NMIM(MY)
STOP WATCH	ES132	SM23148146	18 Jul 2024	SAMM 082	NMIM(MY)

Calibrated By:

Mohd Hafizul

Approved Signatory:

L.H. Seah

This certificate is issued in accordance with the laboratory accreditation requirements of Skim Akreditasi Makmal Malaysia (SAMM) of Standards Malaysia which is a signatory to the ILAC MRA. The measurement results included in this document are traceable to Malaysia national measurement standards maintained by the National Metrology Institute of Malaysia (NMIM). NMIM is a signatory to the CIPM MRA. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the NMIM and other recognised national metrology institutes. The results of calibration performed by Sendi Mahir Sdn Bhd apply to the particular equipment at the time of its test. They do not indicate or imply that Sendi Mahir Sdn Bhd approves, recommends or endorses the manufacturers or suppliers or users of such equipment that Sendi Mahir Sdn Bhd in any way guarantees the equipment's performance after calibration. Test/calibrations marked 'Not SAMM Accredited' in this report/certificate are not included in the SAMM Accreditation Schedule of our laboratory. Opinions and interpretations expressed herein are outside the scope of SAMM accreditation. Copyright of this certificate is owned by the issuing laboratory and may not be reproduced other than in full except with the prior written approval of the Head of the issuing laboratory.