

## BHD. 199501003943 (333138-T)

NO. 6, 8, 10 & 12, JALAN KAPAR 27/89, MEGAH INDUSTRIAL PARK, SEKSYEN 27, 40400 SHAH ALAM, SELANGOR DARUL EHSAN, MALAYSIA.

TEL: 03-5191 7388 FAX: 03-5191 0675 EMAIL: enquiry@sendimahir.com; marketing@sendimahir.com Website: www.sendimahir.com





**SAMM 082** 

## ERTIFICATE OF CALIBRATION

Certificate No.

SM221018723

Date of Issue : 12 Aug 2022

Issued By

Sendi Mahir Sdn Bhd

Page 1 of 2 Pages

Customer

GALAXY AEROSPACE (M) SDN BHD

SUITE 11-14, HELICOPTER CENTRE

MALAYSIA INTERNATIONAL AEROSPACE CENTRE SULTAN ABDUL AZIZ SHAH AIRPORT

47200 SUBANG SELANGOR

Instrument

AC OUTLET ANALYSER AND WRIST STRAP Calibration Date **TESTER** 

: 12 Aug 2022

Manufacturer

RS

Recalibration Date

: 12 Aug 2023

Model/Type

VER-28943

Specified By Customer

Serial No

CTE/264

Remark: The user should be aware that any numbers of factors may cause this instrument to drift out of calibration before the specified calibration interval has expired.

Capacity Resolution Refer to Following Pages

Calibration Environment Condition:

Temperature

: 23.1 to 23.3 °C

Condition Upon

Receiving

Good Physical Condition

Relative Humidity

: 51 to 53 %rh

Condition Upon

Returning

The instrument has been calibrated. The results are as follows. No adjustment done.

Calibration Method:

Internal Calibration Procedure(s) ICPE11

Calibration Venue

Sendi Mahir Sdn Bhd

Measurement

Uncertainty

The reported expanded measurement uncertainty is stated as the standard measurement uncertainty

multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%

and have a coverage factor of k=2 unless stated otherwise.

Reference Standard(s) Used:

Reference Standard Name Serial No Certificate No Due Date Accreditation No Traceability DECADE RESISTANCE BOX E243 SM22981488 01 Mar 2023 **SAMM 082** NPL(UK)

Calibrated By:

Approved Signatory:

L.H. Seah

This certificate is issued in accordance with the laboratory accreditation requirements of Skim Akreditasi Makmal Malaysia (SAMM) of Standards Malaysia which is a signatory to the ILAC MRA. The measurement results included in this document are traceable to Malaysia national measurement standards maintained by the National Metrology Institute of Malaysia (NMIM). NMIM is a signatory to the CIPM MRA. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the NMIM and other recognised national metrology institutes. The results of calibration performed by Sendi Mahir Sdn Bhd apply to the particular equipment at the time of its test. They do not indicate or imply that the Sendi Mahir Sdn Bhd apply to the particular equipment. that Sendi Mahir Sdn Bhd approves, recommends or endorses the manufacturers or suppliers or users of such equipment that Sendi Mahir Sdn Bhd in any way guarantees the equipment's performance after calibration. Test/calibrations marked 'Not SAMM Accredited' in this report/certificate are not included in the SAMM Accreditation Schedule of our laboratory. Opinions and interpretations expressed herein are outside the scope of SAMM accreditation. Copyright of this certificate is owned by the issuing laboratory and may not be reproduced other than in full except with the prior written approval of the Head of the issuing laboratory.

	SERVICEABLE		G	GalaxyAerospace		
	DESCRIPTION	AC OUTLET ANAL	yster with up	PLUG	QTY: 0)	
	PART NO.	776-8315				
	SERIAL NO.	3146008				
	STATUS: OVH/REPAIR / MOD / CAD / TEST / NEW / COMSUMABLE / GSE					
	TSN : NIA		TSO	: NIA		
	SHELF DUE / EXP DATE : 12 PMG UST 3033					
	INSPECTED BY : FROZLIE ADAM			GiN / REF:		
	SIGN:	?.	DATE :	2022	CTE / 264	

- 1

.

.